PIPETTOR CALIBRATION
Annually

PIPETTORS CALIBRATED WITH CLASS A GRADUATE CYCLINDER

PIPETTOR ID: _____ VOLUME OF PIPETTOR _____

DATE	INITIALS	INDIVIDUAL DELIVERIES mL										AVERAGE
		1	2	3	4	5	6	7	8	9	10	

Tolerance ± 2.5%

2ml: 1.95 to 2.05 mL
3ml: 2.925 to 3.075 mL
4ml: 3.9 to 4.1 mL
5ml: 4.875 to 5.125 mL

Comments:

_____

_____