NUTRIENT BROTH QC

DATE	INITIALS	LOT#	TOTAL NUMBER	NUMBER TESTED	VOLUME	pН	SAT/UNSAT

Acceptable Limits: Volume 100 mL \pm 2 mL pH = 6.8 ± 0.2

Volume is checked in 1/25 of each lot pH is checked once per lot